



SECOND DERIVATIVE SPECTROPHOTOMETRIC METHOD FOR THE ESTIMATION OF RIFAMPICIN AND PIPERINE IN THEIR COMBINED DOSAGE FORM

Khamar Jenil C.*, Patel Satish A.

Department of Quality Assurance, S. K. Patel College of Pharmaceutical Education and Research, Ganpat University, Kherva, Mehsana, Gujarat, India

Article Received on: 20/01/12 Revised on: 26/03/12 Approved for publication: 17/04/12

*E-mail: jenilkhamar@yahoo.com

ABSTRACT

The present manuscript describe simple, sensitive, rapid, accurate, precise and cost effective Second derivative spectrophotometric zero crossing point method for the simultaneous determination of Rifampicin and Piperine in combined capsule dosage form. The utility of Second derivative data processing program is its ability to calculate unknown concentration of components of interest in a mixture containing an interfering component. The Second order derivative absorption at 341 nm (zero cross point for Piperine) was used for Rifampicin and 241 nm (zero cross point for Rifampicin) was used for Piperine. The linearity was obtained in the concentration range of 10-60 µg/ml for Rifampicin and 2-20 µg/ml for Piperine. The method was successfully applied to pharmaceutical dosage form because no interference from the capsule excipients was found. The suitability of these methods for the quantitative determination of Rifampicin and Piperine was proved by validation. The proposed methods were found to be simple and sensitive for the routine quality control application of Rifampicin and Piperine in pharmaceutical capsule dosage form. The results of analysis have been validated statistically and by recovery studies.

KEYWORDS: Rifampicin, Derivative spectrophotometric method, Piperine, Drug analysis, Validation, Capsule, Recovery.

INTRODUCTION

Rifampicin (RIFA) is chemically (12Z, 14E, 24E)- (2S, 16S, 17S, 18R, 19R, 20R, 21S, 22R, 23S) - 1,2 -dihydro- 5, 6, 9, 17, 19 -pentahydroxy, 23 -methoxy- 2, 4, 12, 16, 18, 20, 22 heptamethyl -8- (4-methylpiperazin -1 yliminomethyl) -1, 11 - dioxo 2, 7 (epoxypentadeca -1, 11, 13 trienimino) naphtha [2,1-*b*] furan -21-yl acetate.¹ (Figure 1) is a well known Anti-Tuberculosis drug². It is official in IP, BP and USP. IP³, BP⁴ and USP⁵ describe Liquid Chromatography and Visible spectrophotometry method for its estimation. Literature survey reveals HPLC⁶, HPTLC⁷ and Visible Spectrophotometry⁸ methods for determination of RIFA in pharmaceutical dosage forms as well as in biological fluids. Literature survey also reveals spectrophotometric, RP-HPLC^{9,10}, Visible Spectrophotometry^{11,12} and HPTLC¹³ methods for determination of RIFA with other drugs in combination. Piperine (PIPE) is chemically 1-5-(1,3-benzodioxol-5-yl)-1-oxo-2,4-pentadienyl] piperidine¹⁴ (Figure 2) is a natural alkaloid use as Bio enhancer¹⁵. Piperine is official in IP. IP¹⁴ describe liquid chromatography method for its estimation. Literature survey reveals HPLC¹⁶, UV Spectrophotometry¹⁷ and HPTLC^{18,19} method for the determination of PIPE. Literature survey also reveals HPLC^{20,21} method for determination of PIPE with other drugs in combination. The combined dosage forms of RIFA and PIPE along with Isoniazid are available in the market and used as anti tuberculosis drugs. The combination of these two drugs is not official in any pharmacopoeia; hence no official method is available for the simultaneous estimation of RIFA and PIPE in their combined dosage forms. Literature survey does not reveal any simple spectrophotometric method for simultaneous estimation of RIFA and PIPE in combined dosage forms. The present communication describes simple, sensitive, rapid, accurate, precise and cost effective spectrophotometric method based on Second Derivative spectrophotometric method for simultaneous estimation of both drugs in their combined capsule dosage form.

MATERIALS & METHODS

Materials

A shimadzu model 1700 (Japan) double beam UV/Visible spectrophotometer with spectral width of 2 nm, wavelength accuracy of 0.5 nm and a pair of 10 mm matched quartz cell was used to measure absorbance of all the solutions. Spectra were automatically obtained by UV-Probe system software. A Sartorius CP224S analytical balance (Gottingen, Germany), an ultrasonic bath (Frontline FS 4, Mumbai, India) was used in the study. RIFA and PIPE bulk powder was kindly gifted by Cadila Pharmaceuticals Ltd. Ahmedabad, Gujarat, India. The commercial fixed dose combination product was procured from the local market. Methanol (AR Grade, S. D. Fine Chemicals Ltd., Mumbai, India) were used in the study.

Methods

Preparation of Standard Solutions

A 10 mg of standard RIFA and PIPE were weighed and transferred to 100 ml separate volumetric flasks (amber coloured for RIFA) and dissolved in methanol. The flasks were shaken and volumes were made up to mark with methanol to give a solution containing 100 µg/ml each of RIFA and PIPE.

Methodology

The working standard solutions of RIFA and PIPE were prepared separately in methanol having concentration of 10 µg/ml. They were scanned in the wavelength range of 200-800 nm against solvent methanol as blank. The absorption spectra thus obtained were derivatised from first to fourth order. Second order derivative spectrum was selected for the analysis of both the drugs. From the overlain spectra of both the drugs (figure 4) wavelengths selected for quantitation were 341 nm (zero cross point for PIPE) for RIFA and 241 nm (zero cross point for RIFA) for PIPE.

VALIDATION OF THE PROPOSED METHOD

The proposed method was validated according to the International Conference on Harmonization (ICH) guidelines²².

LINEARITY (CALIBRATION CURVE)

Appropriate aliquots from the standard stock solutions of RIFA and PIPE were used to prepare two different sets of dilutions: Series A, and B as follows. Series A consisted of different concentration of RIFA (10-60 µg/ml). Aliquot from the stock solution of RIFA (100 µg/ml) was pipette out in to a series of 10 ml volumetric flask and diluted with methanol to get final concentration in range of 10-60 µg/ml (1, 2, 3, 4, 5 and 6 ml). Series B consisted of varying concentrations of PIPE (2-20 µg/ml). Appropriate volume of the stock solution of PIPE (100 µg/ml) was transferred into a series of 10 ml volumetric flask and the volume was adjusted to the mark with methanol to get final concentration in range of 2-20 µg/ml (0.2, 0.4, 0.8, 1.2, 1.6 and 2.0 ml). The calibration curve were constructed by plotting drug concentration versus the absorbance values of Second derivative spectrum 341 nm for RIFA and 241 nm for PIPE. Statistical data for calibration curves is depicted in Table 1. The concentration of individual drugs present in the mixture was determined from the calibration curves in quantitation mode.

METHOD PRECISION (REPEATABILITY)

The precision of the instrument was checked by repeated scanning and measurement of absorbance of solutions ($n = 6$) for RIFA and PIPE (10 µg/ml for both drugs) without changing the parameter of the proposed spectrophotometry method.

INTERMEDIATE PRECISION (REPRODUCIBILITY)

The intraday and interday precision of the proposed method was determined by analyzing the corresponding responses 3 times on the same day and on 3 different days over a period of 1 week for 3 different concentrations of standard solutions of RIFA and PIPE (10, 20, 40 µg/ml for RIFA and 2, 8, 16 µg/ml for PIPE). The result was reported in terms of relative standard deviation (% RSD).

ACCURACY (RECOVERY STUDY)

The accuracy of the method was determined by calculating the recoveries of RIFA and PIPE by the standard addition method. Known amounts of standard solutions of RIFA and PIPE were added at 50, 100 and 150 % level to prequantified sample solutions of RIFA and PIPE (20 µg/ml for RIFA and 8 µg/ml for PIPE). The amounts of RIFA and PIPE were estimated by applying obtained values to the respective regression line equations.

LIMIT OF DETECTION AND LIMIT OF QUANTIFICATION

The limit of detection (LOD) and the limit of quantification (LOQ) of the drug were derived by calculating the signal-to-noise ratio (S/N, i.e., 3.3 for LOD and 10 for LOQ) using the following equations designated by International Conference on Harmonization (ICH) guidelines.

$$\text{LOD} = 3.3 \times \sigma/S$$

$$\text{LOQ} = 10 \times \sigma/S$$

Where, σ = the standard deviation of the response and
S = slope of the calibration curve.

ANALYSIS OF CAPSULE SAMPLE

Weigh 20 capsules and determine average net content of blend. Remove Isoniazid tablet from blend. Accurately weigh and transfer quantity of capsule contents equivalent to about 200 mg of RIFA and 10 mg of PIPE into 100 ml amber coloured volumetric flask. Add 70 ml of Methanol and sonicate for about 20 minutes. Dilute volume up to mark with Methanol and mix. Take 2 ml aliquot in separate 100 ml amber coloured volumetric flask. Dilute it up to mark with Methanol to get the solution containing 40 µg/ml of RIFA and 2 µg/ml of PIPE. The absorbance of final solution was

recorded at selected wavelengths for determination of RIFA and PIPE. The analysis procedure was repeated three times capsule formulation.

RESULTS AND DISCUSSION

The standard solutions of RIFA and PIPE were scanned separately in the UV range and First-order spectra for RIFA and PIPE were recorded. The first order derivative absorption at 341 nm (zero cross point for PIPE) was used for Rifampicin and 241 nm (zero cross point for RIFA) was used for Piperine. These two wavelengths can be employed for the determination of RIFA and PIPE without any interference from the other drug in their combined formulations.

Linear correlation was obtained between absorbances and concentrations of RIFA and PIPE in the concentration ranges of 10-60 µg/ml and 2-20 µg/ml, with R^2 value 0.999 at both the wavelength respectively. The linearity of the calibration curve was validated by the high values of correlation coefficient of regression. The RSD values of RIFA were found to be 0.31% at 341 nm. The RSD value of PIPE was found to be 0.37% at 241 nm. Relative standard deviation was less than 2 %, which indicates that proposed method is repeatable. The low RSD values of interday (0.32-1.58% for RIFA at 341 nm and 0.14-1.64% for PIPE at 241 nm, respectively) and intraday (0.11-0.69% for RIFA at 341 nm and 0.17-1.21% for PIPE at 241 nm, respectively) variation for RIFA and PIPE, reveal that the proposed method is precise. LOD and LOQ values for RIFA were found to be 1.65 and 5.03 µg/ml at 341 nm, respectively. LOD and LOQ values for PIPE were found to be 0.57 and 1.75 µg/ml at 241 nm, respectively. These data show that method is sensitive for the determination of RIFA and PIPE. The regression analysis data and summary of validation parameters for the proposed method is summarized in Table 1.

The recovery experiment was performed by the standard addition method. The recoveries of RIFA and PIPE were found to 99.63 ± 0.48 and 100.10 ± 0.53 for Rifampicin and Piperine, respectively. The results of recovery studies indicate that the proposed method is highly accurate [Table 2]. The validation parameters are summarized in [Table 1]. The proposed validated spectroscopic method was successfully applied to combined dosage form (Capsule). The results obtained for RIFA and PIPE were comparable with the corresponding label claim percentage [Table 3]. No interference of the excipients with the absorbance of interest appeared; hence the proposed method is applicable for the routine simultaneous estimation of RIFA and PIPE in pharmaceutical dosage forms.

CONCLUSION

The proposed spectrophotometric method was found to be simple, sensitive, accurate and precise for determination of RIFA and PIPE in capsule dosage form. The method utilizes easily available and cheap solvent for analysis of RIFA and PIPE hence the method was also economic for estimation of RIFA and PIPE from capsule dosage form. The common excipients and other additives are usually present in the capsule dosage form do not interfere in the analysis of RIFA and PIPE in method, hence it can be conveniently adopted for routine quality control analysis of the drugs in combined pharmaceutical formulation.

ACKNOWLEDGEMENT

The authors are thankful to Cadila Pharmaceuticals Ltd., Ahmedabad, Gujarat, India for providing gift sample of RIFA and PIPE for research. The authors are highly thankful to Shree S. K. Patel College of Pharmaceutical Education and Research, Ganpat University, Ganpat Vidhyanagar – 384012,

Mehsana, Gujarat, India for providing all the facilities to carry out the work.

REFERENCES

- Maryadele JO Neil. The Merck Index: An Encyclopedia of chemicals, drugs and biologicals. 14th ed., Whitehouse station, New Jersey: Published by Merck Research Laboratories, Division of Merck and Co., Inc; 2006: 1417.
- Sweetman S.C. The Martindale: The Complete Drug Reference. 35th ed. Pharmaceutical Press. London, UK: 2007. p. 290.
- Indian Pharmacopoeia, Vol. III, New Delhi, The Controller Publication, Govt. of India; 2010: 2054-2065.
- British Pharmacopoeia, Vol. II, London, The British Pharmacopoeia Commission; 2010: 1844, 3063.
- The United State Pharmacopoeia, USP28 NF23, Rockville MD, United State Pharmacopoeial Convention, Inc; 2005: 3501 – 3507.
- Panchagnula R, Sood A, Sharda N, Kaur K, Kaul CL, Determination of rifampicin and its main metabolite in plasma and urine in presence of pyrazinamide and isoniazid by HPLC method, Journal of Pharmaceutical and Biomedical Analysis January 1999, 18(6), 1013–1020
- Shishoo CJ, Shah SA, Rathod IS, Savale SS, Vora MJ, Impaired bioavailability of rifampicin in presence of isoniazid from fixed dose combination (FDC) formulation, International Journal of Pharmaceutics, October 2001; 228(1-2); 53–67.
- Mariappan TT, Jindal KC, Singh S, Overestimation of rifampicin during colorimetric analysis of anti-tuberculosis products containing isoniazid due to formation of isonicotinyl hydrazone, Journal of Pharmaceutical and Biomedical Analysis, November 2004; 36(3); 905–908
- Khuhawar MY and Rind FMA, High performance liquid chromatographic determination of isoniazid, pyrazinamide and rifampicin in pharmaceutical preparations, Pakistan journal of pharmaceutical sciences; July 1998; 18(2); 49-54.
- Calleri E, Lorenzi EDe, Furlanetto S, Massolini G, Caccialanza G, Validation of a RP-LC method for the simultaneous determination of isoniazid, pyrazinamide and rifampicin in a pharmaceutical formulation, Journal of Pharmaceutical and Biomedical Analysis, August 2002; 29(6); 1089–1096
- Manna A, Ghosh I, Datta S, Ghosh PK, Ghosh LK, Gupta BK, Simultaneous Estimation Of Rifampicin And Isoniazid In Combined Dosage Forms, Indian journal of pharmaceutical science; 2000; 62(3); 185-186
- Goyal P, Pandey S, Udupa N, Simultaneous Spectrophotometric Estimation Of Isoniazid And Rifampicin From Combined Dosage Forms Indian journal of Pharmaceutical science; 2002; 64(1); 76-78
- Ali J, Ali N, Sultana Y, Baboota S and Faiyaz S, Development and validation Of a stability-indicating hptlc method For analysis of antitubercular drugs, acta chromatographica, no. 18, 2007, 168-179.
- Indian Pharmacopoeia, Vol. III, New Delhi, The Controller Publication, Govt. of India; 2010: 2522, 2530-31.
- Atal CK, Dubey RK, Singh J. "Biochemical basis of enhanced drug bioavailability by piperine: evidence that piperine is a potent inhibitor of drug metabolism". J. Pharmacol. Exp. Ther., January 1985; 232 (1); 258–62.
- Wood AB, Barrow ML, James DJ, Piperine determination in pepper (*Piper nigrum* L.) and its oleoresins—a reversed-phase high-performance liquid chromatographic method, Issue Flavour and Fragrance Journal, June 1988; 3(2); 55–64.
- Gupta V and Jain UK., Estimation of Piperine By UV-Spectrophotometric Method In Herbal Formulation, Pippli Churna, International Journal of Research in Pharmaceutical and Biomedical Sciences, Apr – Jun 2011; 2 (2); 550-553
- Hamrapurkar PD, Jadhav K and Zine S, Quantitative Estimation of Piperine in *Piper nigrum* and *Piper longum* Using High Performance Thin Layer Chromatography Journal of Applied Pharmaceutical Science 2011; 01 (03); 117-120
- Shanmugasundaram P, Maheswari R, and Vijayaandhi M, Quantitative estimation of piperine in herbal cough syrup by hptlc method, Rasayan J. Chem, 2008; 1(2); 212-217
- Nagappan KV, Meyyanathan SN, Raja RB and Kannan E, A Liquid Chromatography Method for the Simultaneous Determination of Curcumin and Piperine in Food Products Using Diode Array Detection, Asian J. Research Chem.; April.-June, 2009, 2(2): 115-118
- Kamal YT, Mohammed Musthaba S, Singh M, Parveen R, Ahmad S, Baboota S, Ali I, Siddiqui KM, Arif Zaidi SM.; Development and validation of HPLC method for simultaneous estimation of piperine and guggulsterones in compound Unani formulation (tablets) and a nanoreservoir system. Biomedical Chromatography, Dec-2011, DOI: 10.1002/bmc.2676.
- The International Conference on Harmonization, Q2 (R1), Validation of Analytical Procedure: Text and Methodology: 2005.




Figure 1: Structure of Piperine




Figure 2: Structure of Rifampicin




Figure 3: Overlain zero-order absorption spectra of RIFA and PIPE in methanol




Figure 4: Overlain Second order Derivative absorption spectra of RIFA and PIPE in methanol

Table 1: Regression Analysis Data and Summary of Validation Parameters for RIFA and PIPE by Second Derivative Spectrophotometric Method

| Parameters | RIFA | PIPE |
|--|----------------------|----------------------|
| Wavelength (nm) | 341 | 241 |
| Beer's law limit (µg/ml) | 10-60 | 2-20 |
| Regression equation (y = a + bc) | y = 0.0001x + 0.0002 | y = 0.0002x - 0.0002 |
| Slope (b) | 0.00013 | 0.00017 |
| Intercept (a) | + 0.0002 | - 0.0002 |
| Correlation coefficient (r ²) | 0.999 | 0.999 |
| LOD ^a (µg/ml) | 1.65 | 0.57 |
| LOQ ^b (µg/ml) | 5.03 | 1.75 |
| Repeatability (% RSD ^c , n=6) | 0.31 | 0.37 |
| Precision (%RSD, n = 3) | | |
| Interday | 0.32-1.58 | 0.14-1.64 |
| Intraday | 0.11-0.69 | 0.17-1.21 |
| Accuracy ± S.D. ^d (%Recovery, n= 5) | 99.63 ± 0.48 | 100.10 ± 0.53 |

^aLOD = Limit of detection, ^bLOQ = Limit of quantification, ^cRSD = Relative standard deviation. ^dS. D. = Standard deviation

Table 2: Recovery Data of RIFA and PIPE by Spectrophotometric Method

| Drug | Amount taken (µg/ml) | Amount added (%) | %Recovery ± S. D. (n=5) |
|------|----------------------|------------------|-------------------------|
| RIFA | 20 | 50 | 100.18 ± 0.65 |
| | 20 | 100 | 99.18 ± 0.32 |
| | 20 | 150 | 99.54 ± 0.48 |
| PIPE | 8 | 50 | 100.52 ± 0.76 |
| | 8 | 100 | 99.45 ± 0.26 |
| | 8 | 150 | 100.34 ± 0.58 |

Table 3: Analysis of RIFA and PIPE by Spectrophotometric Method

| Capsule | Label Claim (mg) | | Amount Found (mg) | | % Label Claim ± S.D. (n=6) | |
|---------|------------------|------|-------------------|------|----------------------------|--------------|
| | RIFA | PIPE | RIFA | PIPE | RIFA | PIPE |
| I | 200 | 10 | 199.76 | 9.96 | 99.88 ± 0.22 | 99.69 ± 0.28 |

Source of support: Nil, Conflict of interest: None Declared